## OFFICE OF ATTORNEY GENERAL CRIME LABORATORY DIVISION

Toxicology Section/Breath Program Chain of Custody Worksheet

Worksheet: BrW-012

## CHAIN OF CUSTODY BREATH ALCOHOL INSTRUMENT

Instrument: 8000	Serial Number: 80-007097			
Agency being assigned instrument: NDOAG - Crime Lab				
Date 1 Nov 18				
Transferred from:	<u>Transferred to</u> :			
Name (print) Inc .	Name (print) Roberta Grieges - Nimmo			
Signature	Signature Hotal Sugar nimm			
Agency CMT, Inc.	Agency NDOAG-Crime Lab			
Reason for transfer/remarks:  CMI, Inc. performed repair. Repositioned battery wiving harnest				
Or Shipped Via:				
USPS UPS _X_ FedEx				
Date Sending Instrument://				



DATE RCVD VIA

316 E. 9<sup>th</sup> St., Owensboro KY 42303 Phone 866-835-0690 Fax: 270-685-6268

## SERVICE WORK ORDER

BILL TO:	SHIP TO: NO (King (ab				
ATTN:	ATTN: Charles Eder				
	2641 E. Main X10.				
CITY STATE ZIP	CITY. STATE ZIP				
CUSTOMER NO.	PHONE 701-328-6159				
PHONE	FAX 701-328-4185				
MODEL # 8000 SERIAL # 0-007097	00 2480 KD BILL CODE 3				
BT, an wax, bustoner					
DESCRIPTION OF PROBLEM BALLY WON'T CHANGE					
Infrared / Fuel Cell PRELIMINARY TEST INFORMATION  BATTERY- EXT	PARTS USED				
External battery won't change	QTY PART NO. DESCRIPTION				
	Yes /				
WORK PERFORMED:					
Repositioned battery wi	in a				
harness.					
Miscellaneous PRELIMINARY TEST INFORMATION					
WORK PERFORMED:					
Por					
REPAIR CAL ADJ. CALIBRATON FINAL					

REPAIR	CAL. ADJ.	CALIBRATON	FINAL
½ HOURS	HOURS	HOURS	HOURS
e e e e e e e e e e e e e e e e e e e	man ginas junta minde etilet film til Erminister et rette til ste på til ste film til ste film til ste film ti		(/
DATE	DATE	DATE	DATE
Oct 23,14	Pay management in a management in a service consistency of the service consistency of the service consistency	((1)) N	Cato M

SPECIAL SHIPPING INSTRUCTIONS